



DR. BHUBANESWAR BOROOAH CANCER INSTITUTE
PROCUREMENT DEPARTMENT

ITEM DESCRIPTION : **CORRIGENDUM FOR “SYRINGE PUMP”**
GEM BID NO : **GEM/2024/B/5107342**

Sr. No	Original specification	Revised Technical Specification
1.	Operational Requirement The syringe pump should be programmable, user friendly, safe to use and should have battery backup and comprehensive alarm system. This should be able to integrate in the HIS.	Read as: Operational Requirement The syringe pump should be programmable, user friendly, safe to use and should have battery backup and comprehensive alarm system. This should be able to integrate in the HIS, should be complied with HL7 protocol. Supportive document for HL7 compatibility to be provided.

Dr.
Director
22/7/24
Dr. B. Borooah Cancer Institute